

## Examination Glove Nitrile Top Pro



EN ISO 374-5:2016 EN ISO 374-1:2016  
+A1:2018 Type C



VIRUS



KT

<b>Description</b>	Tear-resistant Nitrile-Examination glove Length: 24 cm AQL: 1,5, non-sterile Weight Size M: 7,3 g/piece +/-0,5g
<b>Material</b>	Nitrile
<b>Application areas</b>	Medicine, agriculture, cleaning, light chemical protection
<b>Produktvorteile</b>	Durable and tear-proof ambidextrous food safe latex- and powderfree free from allergenic thiurams, thiazoles and latex proteins
<b>Single packaging</b>	100 pieces per dispenser box
<b>Standards</b>	KAT III, EN ISO 374-1:2016+A1:2018 Type C, EN ISO 374-5:2016 (EU-Regulation 2016/425) EN 455/1-4 (EU-Regulation 2017/745 Class 1 medical device)

### Logistical Data:

Item no.	Size	Barcode	Minimum order quantity	Packaging Unit	Pal
153070	S / 6,5-7	4018653302101	1 Pack	1/10	360
153071	M / 7,5-8	4018653302118	1 Pack	1/10	360
153072	L / 8,5-9	4018653302125	1 Pack	1/10	360
153073	XL / 9,5-10	4018653302132	1 Pack	1/10	360
153074	XXL / 10,5-11	4018653302149	1 Pack	1/10	360

**Certification:**

**A) Chemical resistance according to EN ISO 374-1:2016+A1:2018, Type C, EN 374-4:2013**

ID Code	Substance	CAS-Nr.	Class	Performance level*	Degradation
K	40 % Sodium Hydroxide	1310-73-2	Inorganic base	6	-10,1 %
O	25 % Ammonium Hydroxide	1336-21-6	Organic base	1	60,8 %
T	37 % Formaldehyde	50-00-0	Aldehyde	6	45,4 %

\*legend:

Performance level	1	2	3	4	5	6
Breakthrough time (min)	>10	>30	>60	>120	>240	>480

**B) Microbial resistance according to EN ISO 374-5:2016**

Protection against bacteria and fungal spores: pass

Protection against virus: pass

**C) Medical device according to EN 455/1-4**

SRN:	DE-MF-000007708
Basic UDI-DI:	40186530074C